



Certificate of Traceability

Level I Laboratory Certification

This certifies that LabCorp Central Laboratory, using Roche Cobas c503 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Cobas c503	Calibrator Lot: 67182401, 67182401	Application: (HbA1c/Hb)*91.5+2.15
Reagent Lot: 718887, 732837, 699282	Calibrator Assigned Values: 0.156, 0.360, 0.686, 0.939, 1.29, 1.82 mmol/L HbA1c; 1.48, 7.97 mmol/L Hb	

Date of Certification: January 1, 2024

Certification Expires: January 1, 2025

David Seals

NGSP Steering Committee Chair

Randee Little PhD.

NGSP Network Coordinator

Sharon Amulfo

SRL director/ supervisor